

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

APR 20 2010

Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130660</u>	
Name <u>Garrett Willis</u>		ID# <u>5560</u>	Date <u>4/18/10</u>
A	Agency <u>AST</u>		Phone # <u>262-4453</u>
Instrument Location <u>Cooper Landing</u>			
Alco S/N <u>57345</u>		Target Value <u>.082</u>	High Pressure <u>800</u>
B	Alco Test Value Average <u>.085</u>		<u>.085</u>
	1 st Alco	2 nd Alco	
Signature <u>Garrett Willis</u>		<u>COB</u> <u>4/20/10</u>	
(OVER)			

I, Nita J. Bolz, after being first duly sworn, depose and state as follows (in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 20th day of April, 2010.

Carolyn M. Noland (Notary Seal Stamp)
Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130660 ✓

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130660 ✓

APRIL 16, 2010 ✓
CDB 4/26/10

OPERATOR'S NAME:
WILLIS/GARRETT/R
OPERATOR'S NUMBER: 5560
SUBJECT'S LAST NAME:
VOC
SUBJECT'S FIRST NAME/MI :
VOC
O.L. #: VOC
DEPT/AGENCY: C000
CASE/REPORT: VOC
TEST TYPE: V
ALCO TARGET VALUE: .082 ✓
ALCO S/N: 57345 ✓

--- BREATH ANALYSIS ---

.082 ADJUSTED FOR 29.60 in
ALCO TARGET .081 09:07
BLANK TEST .000 09:07
INTERNAL STANDARD VERIFIED 09:08
ALCO TV 29.59 in .085 ✓ 09:08
BLANK TEST .000 09:09
SUBJECT SAMPLE CDB 4/26/10 .000 09:09
BLANK TEST .000 09:10
ALCO TV 29.59 in .085 ✓ 09:10
BLANK TEST .000 09:11

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130660 ✓

APRIL 16, 2010 ✓
TIME 09:04

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.60 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

! " # \$ % & ' () * + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G
H I J K L M N O P Q R S T U V W X Y Z [\] ^ _ ` a b c d e f g h i j k l m n o
p q r s t u v w x y z { | } ~